



# CERTIFICATE OF ACCREDITATION

## ANSI-ASQ National Accreditation Board

500 Montgomery Street, Suite 625, Alexandria, VA 22314, 877-344-3044

This is to certify that

**Morgan Schaffer Ltd**  
**8300 Saint-Patrick Street, Suite 150**  
**LaSalle, QC H8N 2H1**

has been assessed by ANAB  
and meets the requirements of international standard

## ISO Guide 34:2009

while demonstrating technical competence in the field of

## Reference Material Producer

Refer to the accompanying Scope of Accreditation for information regarding the types of materials to which this accreditation applies.

AR-2126

Certificate Number

  
ANAB Approval

Certificate Valid: 07/27/2017-03/08/2019  
Version No. 002 Issued: 07/27/2017





### SCOPE OF ACCREDITATION TO ISO GUIDE 34:2009

#### Morgan Schaffer Ltd

8300 Saint-Patrick Street, Suite 150

LaSalle, QC H8N 1H1

Contact: Denis Lafrance Phone: 514-739-1967

Email: dlafrance@morganschaffer.com Website: www.morganschaffer.com

### REFERENCE MATERIAL PRODUCER

Valid to: **March 8, 2019**

Certificate Number: **AR-2126**

Sub-Category of Reference Material	ILAC RM Category	Class or Type of Reference Materials Produced (Include Range Where Applicable)	Methods or Techniques Used in the RMP Laboratory (if Appropriate)
Gases (Certified Reference Material)	A7.1	Gas mixture in dielectric fluids (5, 10, 50, 100, 500 and 1000 ppm ( $\mu\text{g/g}$ ) of $\text{H}_2$ , $\text{CH}_4$ , $\text{C}_2\text{H}_6$ , $\text{C}_2\text{H}_4$ , $\text{C}_2\text{H}_2$ , $\text{CO}_2$ , $\text{O}_2$ , $\text{N}_2$ , and $\text{CO}$ )	Gas Chromatography
Moisture (Certified Reference Material)	A3.5	Moisture in dielectric fluids 15 and 30 ppm ( $\mu\text{g/g}$ )	Karl Fisher Titrator

Notes:

1. Please contact the RMP organization for more information on CRM uncertainty values, Ucrm values, and other specific lot values. Some of this information may also be available on the RMP's website.
2. This scope is formatted as part of a single document including Certificate of Accreditation No. AR-2126.



Vice President

